



AURELIA® BOLD

5mil, Nitrile Powder Free Examination Gloves

Regulations	PPE Reugulation (EU) 2016/425 Medical Device Regulation (EU) 2017/745, 93/42/EEC
Quality Standards	ISO 9001:2015 Quality Management Systems ISO 13485: 2016 Medical Devices Quality Management Systems
Harmonised Standards	BS EN ISO 374 - 1,2,4,5 Protective gloves against dangerous chemicals and microorganisms. BS EN 455 1-4 Medical gloves for single use. EN 420: 2003 + A1: 2009 Protective gloves. General requirements and test methods. BS EN 1186: 2002. Materials and articles in contact with foodstuffs.

AURELIA® GLOVES

A SUPERMAX HEALTHCARE BRAND

sales@supermax.co.uk
+441733 615215

Supermax Healthcare Ltd,
Unit 12-16 Titan Drive
Fengate, Peterborough
PE1 5XN

aureliagloves.co.uk



10 x 100



UK
CA
0120



S M L XL XXL

73996 73997 73998 73999 73990

Sizes & Dimensions

Single wall thickness

Physical Properties

XS 70+10 mm	min. 240 mm
S 80+10 mm	min. 240 mm
M 95+10 mm	min. 240 mm
L 110+10 mm	min. 240 mm
XL > 110 mm	min. 240 mm



- ① min. 0.13 mm
- ② min. 0.11 mm
- ③ min. 0.08 mm

	Unaged	Aged
Tensile Strength	Min. 14.0 MPa	Min. 14.0 MPa
Elongation	Min. 500%	Min. 400%
Strength at Break	Min. 6N	Min. 6N



Nitrile



Powder Free



5 Years



Micro



5.5g +/- 0.2g



Black

